K-MANO U

Pressure gauges (CrNi steel type / connection radial on bottom)

HANSA-FLEX

Caractéristiques

Type	232.50
Fabrication	Bourdon-tube pressure gauge, CrNi steel type, all-stainless steel
Application	For gaseous or liquid media which do not corrode copper alloy,do not have high viscosity and do not crystallise
Classe de qualité	1.6 (Ø 63 mm), 1.0 (Ø 100 mm and Ø 160 mm)
Température de fluide	max. +200 °C
Température ambiante	-40 °C to +60 °C
Boîtier	CrNi steel
Lumière	Laminated safety glass Ø 63 = Polycarbonate



Remarque

Autres informations sur demande.

Article

Désignation	Plage de mesure	Ø (mm)	Raccord
K- 07 20 06 01	-1 / 0.0 bar	63,0	G 1/4"
K- 07 20 06 02	0 - 2.5 bar	63,0	G 1/4"
K- 07 20 06 03	0 - 4.0 bar	63,0	G 1/4"
K- 07 20 06 04	0 - 6.0 bar	63,0	G 1/4"
K- 07 20 06 05	0 - 10.0 bar	63,0	G 1/4"
K- 07 20 06 06	0 - 16.0 bar	63,0	G 1/4"
K- 07 20 06 07	0 - 25.0 bar	63,0	G 1/4"
K- 07 20 06 08	0 - 40.0 bar	63,0	G 1/4"
K- 07 20 06 09	0 - 100.0 bar	63,0	G 1/4"
K- 07 20 00 36	-1 / 0.0 bar	100,0	G 1/2"
K- 07 20 00 37	-1 / +0.6 bar	100,0	G 1/2"
K- 07 20 00 38	-1 / +1.5 bar	100,0	G 1/2"
K- 07 20 00 39	-1 / +3.0 bar	100,0	G 1/2"
K- 07 20 00 40	-1 / +5.0 bar	100,0	G 1/2"
K- 07 20 00 41	-1 / +9.0 bar	100,0	G 1/2"
K- 07 20 00 42	-1 / +15.0 bar	100,0	G 1/2"
K- 07 20 00 43	0 - 1.0 bar	100,0	G 1/2"
K- 07 20 00 44	0 - 1.6 bar	100,0	G 1/2"
K- 07 20 00 45	0 - 2.5 bar	100,0	G 1/2"
K- 07 20 00 46	0 - 4.0 bar	100,0	G 1/2"
K- 07 20 00 47	0 - 6.0 bar	100,0	G 1/2"
K- 07 20 00 48	0 - 10.0 bar	100,0	G 1/2"
K- 07 20 00 49	0 - 16.0 bar	100,0	G 1/2"
K- 07 20 00 50	0 - 25.0 bar	100,0	G 1/2"
K- 07 20 00 51	0 - 40.0 bar	100,0	G 1/2"
K- 07 20 00 52	0 - 60.0 bar	100,0	G 1/2"
K- 07 20 03 10	-1 / 0.0 bar	160,0	G 1/2"
K- 07 20 03 11	-1 / +0.6 bar	160,0	G 1/2"
K- 07 20 03 12	0 - 1.0 bar	160,0	G 1/2"
K- 07 20 03 13	0 - 10.0 bar	160,0	G 1/2"
K- 07 20 03 14	0 - 16.0 bar	160,0	G 1/2"
K- 07 20 03 15	0 - 25.0 bar	160,0	G 1/2"

En dépit du soin apporté aux vérifications, nous ne pouvons garantir une absence totale d'erreur et ne saurions être tenus responsables des indications contenues.